

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

This discussion paper is/has been under review for the journal Atmospheric Measurement Techniques (AMT). Please refer to the corresponding final paper in AMT if available.

A fast and precise chemiluminescence ozone detector for eddy flux and airborne application

**A. Zahn¹, J. Weppner¹, H. Widmann¹, K. Schlote-Holubek², B. Burger²,
T. Kühner², and H. Franke³**

¹Karlsruhe Institute of Technology (KIT), Institute for Meteorology and Climate Research (IMK), Germany

²Karlsruhe Institute of Technology (KIT), Institute for Data Processing and Electronics (IPE), Germany

³Enviscope GmbH, Frankfurt, Germany

Received: 14 September 2011 – Accepted: 18 October 2011 – Published: 28 October 2011

Correspondence to: A. Zahn (andreas.zahn@kit.edu)

Published by Copernicus Publications on behalf of the European Geosciences Union.

[Title Page](#)

[Abstract](#) [Introduction](#)

[Conclusions](#) [References](#)

[Tables](#) [Figures](#)

[⏪](#) [⏩](#)

[◀](#) [▶](#)

[Back](#) [Close](#)

[Full Screen / Esc](#)

[Printer-friendly Version](#)

[Interactive Discussion](#)

Abstract

A commercially available dry chemiluminescence (CI) instrument for fast and precise measurement of ozone (O_3) is specified. The sensitivity is ~ 9000 counts s^{-1} per ppbv of ozone. Its precision is entirely determined by the number of photons reaching the detector (being a photomultiplier), i.e. is quantum-noise limited. The relative precision ($\Delta O_3/O_3$ in %) thus follows Poisson statistics and scales with the square root of the measurement frequency f and with the inverse O_3 mixing ratio: $\Delta O_3/O_3 \propto f^{0.5} \cdot O_3^{-0.5}$. At typical O_3 mixing ratios between 10 and 100 ppbv, the precision is 0.3–1.0% at $f = 10$ Hz. The maximum measurement frequency is 50 Hz. The mechanical and electronic set-up as well as the instrument performance is described. Recommendations on the adequate inlet tube configuration (inlet tube length, sampling flow) and on the way of calibration at stationary ground-based platforms and onboard aircraft are given.

1 Introduction

Ozone (O_3) is both chemically and radiatively one of the most important trace gases in the atmosphere. It is the major precursor of the hydroxyl radical (OH), the most efficient detergent in the atmosphere (Levy, 1971), and it forms the stratospheric ozone layer shielding the Earth's surface from harmful UV sunlight (WMO, 2007). Tropospheric ozone is also a potent natural and anthropogenic greenhouse gas (IPCC, 2007). Ozone thus belongs to the most frequently measured atmospheric trace gases.

There are various types of in-situ ozone instruments in use which are based on electro-chemical, spectroscopic and chemiluminescence (CI) techniques (Weinheimer 2006; Li et al., 2008; Vömel and Diaz, 2010). The largest advantage of CI techniques is their fast response time in the range of 1 to 20 Hz. One distinguishes between gas-phase CI and heterogeneous or dry CI techniques.

Gas-phase CI instruments are more complex and are based on the reaction of ozone with either ethene or nitrogen oxide (Ray et al., 1986; Ridley et al., 1992; Hegglin et al., 2006; Williams et al., 2006; Bariteau et al., 2010).

AMTD

4, 6539–6575, 2011

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Dry CI techniques are quite simple and use organic dyes (such as luminol, rhodamine-B, eosin-Y, or coumarin adsorbed on solid surfaces/discs) whose reactions with ozone yields to the emission of photons (usually in the visible wavelength range) that are detected with a photomultiplier. The first instrument was described by Re-

gener (1960). For an excellent review, especially of the early work in the 1960s to 1980s, see Weinheimer (2006) and Muller et al. (2010, and references therein).

Important to note is the fact that dry CI techniques are not absolute, i.e., they need frequent calibration in periods of 1 to 60 min, depending on the scientific requirements, the sensor disc that is used, and ambient conditions (e.g. water mixing ratio). See Sect. 6 for more discussion.

To our knowledge, four types of dry CI ozone instruments have been used during the last ~20 yr: (1) a commercial, battery-powered instrument (Drummond et al., 1991) uses eosin-Y as dye (Ray et al., 1986). (2) A device manufactured by the company GFAS is based on the instrument described by Schurath et al. (1991) and Guesten et al. (1992). It uses Coumarin-47 as dye and was sold until the year ~2000. Modified GFAS devices were also deployed onboard light-weight aircraft (McKendry et al., 1998) and onboard the high-flying research aircraft Egrett (Bradshaw et al., 2002). Moreover, two clones of the GFAS sensor are in use, i.e. the Rapid Ozone Flux Instrument (ROFI) described in Coyle (2005) and Muller et al. (2010), and the 10 Hz-analyzer built by Jim Womack at NOAA-ATDD (Kurpius et al., 2002). (3) FOZAN I and II uses Coumarin-307 and was developed for the deployment on board the high-flying Russian aircraft Geophysica (Yushkov et al., 1999; Ulanovsky et al., 2001). (4) A miniaturized version of the GFAS sonde that is regularly deployed on board the CARIBIC passenger aircraft (Zahn et al., 2002; Brenninkmeijer et al., 2007) which finally resulted in the commercially available instrument described here.

Indeed, fast measurement of ozone is a subject within diverse research fields such as in atmospheric chemistry and bio-chemistry. To our knowledge, however, a fast off-the-shelf instrument is commercially not available at present.

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

This situation comprises the background for the co-operation between KIT and Enviscope GmbH, aiming for optimizing the available CARIBIC sonde in terms of performance (high precision and response time) and usability (high robustness and reliability, simple installation and maintenance). The instrument development actually took advantage of the long development history of the CI technique at our institute (Güsten et al., 1992) and the extensive experiences collected on board the CARIBIC aircraft since 1997.

In this paper we document the mechanical and electrical set-up and the performance of this new ozone instrument. It is however not the objective to describe the properties of CI sensors discs, which is e.g. documented by Sahand (1989), Speusser (1989), and Schurath et al. (1991).

2 Mechanical setup

The CARIBIC device is optimized for the use on board aircraft, i.e. is very light-weight and compact. In constructing a commercial device, however, also functional as well as design specifications had to be met. The mechanical setup of the original CARIBIC device has thus been optimized in terms of robustness and facile handling by the company Enviscope.

Figure 1 shows a schematic of the instrument. Sample air enters the device via a PFA inlet fitting and reaches after a ~ 30 cm (black and opaque) $1/4''$ PTFE tube the sensor disc where CI reactions with ozone emit light (photons). The entrance window of the photomultiplier detecting the photons emitted by the sensor disc is placed directly opposite of it, in a distance of 4 mm. All parts within the sample air channel are made of chemically inert materials such as PTFE, PFA and PEEK. To ensure complete light tightness, the entire system was capsulated by an aluminum housing. To avoid corona effects at the photomultiplier, a 1 cm tall copper hull connected to the cathode high voltage was placed around the photomultiplier window.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

In standard field configuration, sample air is driven through the instrument by a high-efficient van (Micronel D340Q) at the air outlet. If connecting longer sampling tubes, the van can easily be removed to connect an external pump in order to maintain short response times of ≤ 100 ms (see Sects. 4.5 and 6.2).

The sensor disc (diameter: 10 mm) is fixed on top of a small cylindrical holder. To prevent temperature-driven changes of the sensitivity of the sensor disc of typically 2 % per Kelvin (Schurath et al., 1991), the sensor disc holder is temperature stabilized to ~ 38 °C by an integrated temperature controller. The sensor disc holder can easily be removed by inserting a quick-release knob and turning it by 90°. This mechanism allows exchange of the sensor disc within a few seconds and makes it possible to maintain the instrument even under adverse operational conditions, e.g. on high towers. To avoid damage of the photomultiplier, its high-voltage supply is automatically switched off while exchanging the sensor disc. Pressure tightness is guaranteed by O-ring seals allowing the deployment of the instrument onboard aircraft, too.

3 Electronics and operation

The electronics responsible for instrument control, i.e. parameter settings, data acquisition and data pre-treatment was developed at the Institute for data processing and electronics (IPE) at the Karlsruhe Institute of Technology (KIT).

As outlined in Sect. 2 and further specified in Fig. 2, main electric parts are housed in two aluminum compartments:

- (a) the photomultiplier together with its high voltage supply. We choose a small channel photomultiplier (CPM) of PerkinElmer (type C944), as it shows thermionic noise only, in contrast to usual PM where also dynode noise and leakage current significantly contribute to the total noise.
- (b) the main circuit board encompassing power supply, frontend electronics, analog-to-digital converter (ADC), digital-to-analog converter (DAC), and microcontroller.

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Considerable effort was undertaken to keep these components on the one hand as compact as possible and on the other hand extremely precise and EMI-safe.

The CPM is supplied by a high-voltage of typically 1700 V which generates a current amplification of $\sim 4.5 \cdot 10^5$ (see CPM datasheet). This amplification leads to relevant CPM currents I_{CPM} of ~ 1 nA per ppbv of O_3 (1 ppbv = 10^{-9} mol mol $^{-1}$), see Sect. 5. The CPM and the high voltage supply are controlled by the microcontroller via a 16-bit DAC.

In a following high-accuracy low-noise transimpedance amplifier (frontend module) the CPM current I_{CPM} is amplified with a total amplification resistance of 2 M Ω . The resulting voltage $U_{2\text{M}\Omega}$ is digitized by a low noise 24-bit delta-sigma analog-to-digital converter (ADC, ADS1256 from Burr-Brown, Texas Instruments) and multiplied with selectable programmable gain amplification (PGA). A maximum input voltage $U_{2\text{M}\Omega}$ of 5 V corresponds to 23-bit (8 388 608 counts). The ADC integrates over a measurement cycle of 20 ms and passes this 23-bit information (at 50 Hz) via a serial-peripheral-interface (SPI) to the microcontroller. The 1- σ noise (at 5 Hz) is 6 counts or 4 μV , respectively.

The entire system is controlled using a microcontroller from NXP (LPC2138, Philips) which is a 16/32-bit controller equipped with a CPU of Arm (ARM7TDMI-S). This microcontroller includes various 32-bit timers, dual 10-bit 8-channel ADCs, a 10-bit digital-to-analog converter (DAC), pulse-wide-modulation (PWM) channels and 47 general purpose input/output (GPIO) lines. A 128-bit wide interface/accelerator enables high-speed 60 MHz operation.

There are two possibilities to communicate with the microcontroller and to set the system parameters: (a) via a serial RS232 (COM) interface (via a simple ASCII protocol), and (b) via three push buttons and an organic LED-display located at the instrument front plate. The following system parameters can be adjusted:

- High-voltage for the channel photomultiplier U_{CPM} (between 0 and 3000 V)
- Measurement frequency f (1, 10, 20, 50 Hz)
- Programmable amplification (PGA) of the ADC (1, 2, 4, 8, 16, 32, and 64)

- Adjustment of the RS-232 dataset (e.g. to send also the measured CPM high-voltage and the actual temperature of the electronics)

The set measurement frequency defines the number of 20 ms measurements used to calculate the mean value and its standard deviation, e.g. at $f = 10$ Hz five 20 ms measurements are considered. The time, signal mean value, its standard deviation, and (depending on the set protocol) further parameters are sent with the set measurement frequency via RS-232 (baud rate 38 400). These data are also and automatically stored onto a USB stick, if inserted in the USB port at the front plate.

Besides setting parameters the LED-display depicts the measured voltage $U_{2M\Omega}$ (multiplied with the set PGA value). After entering the actual ozone mixing ratio, the microcontroller calculates the actual sensor sensitivity and the display also illustrates the ozone mixing ratio (simply by applying the just measured sensor sensitivity). The analog output range is specified as 0–5 V.

4 Instrument characterization

In this chapter we describe the instrument properties, in particular the measurement noise or precision, respectively, in dependence on diverse parameters such as photo-multiplier voltage, O_3 mixing ratio, measurement frequency, and sample air flow.

As detector discs ~10 yr old discs manufactured by the company GFAS are used. To our knowledge, the exact preparation is not documented, but is assumed to have been carried out as documented by Sahand (1989) and Speuser (1989). That is, 5 g Coumarin-47 and 16 g gallic acid is dissolved in 1 l methanol and is thereafter 1:1 mixed with a silicon-impregnation agent. After pipetting this emulsion on silica gel discs and drying for 1.5 h at 60 °C, a further hydrophobic fluid is added to reduce the humidity dependence.

The sample air used for all tests described here was produced by an O_3 generator SYCOS KT-O3M (Ansyco, Germany) which provides O_3 mixing ratios of up to 800 ppbv in dry air.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



4.1 Precision as function of photomultiplier amplification

The channel photomultiplier (CPM) does not work in a single photon counting mode, but in continuous mode where the CPM output current is measured with an analog-digital converter (ADC) as DC voltage. The CPM output current scales with the high-voltage (HV) applied to the CPM, i.e. with the HV dependent current amplification given in the CPM datasheet. First, we determine the optimal operating range of the CPM.

In Fig. 3, the absolute signal and the relative noise of the instrument are shown at a typical ambient O₃ mixing ratio of 45 ppbv and a measurement frequency of 10 Hz in dependence of the high-voltage HV applied to the CPM.

The signal increases from ~3000 counts (1.81 mV) at HV = 1360 V to ~1 500 000 counts (904 mV) at HV = 1900 V, i.e. by a factor of ~500. Interestingly, the relative noise (here the standard deviation of the 10 Hz measurements over 30 s, i.e. 300 data points) remains constant at (0.55 ± 0.05)%. This observation elucidates the following instrument properties:

- The electrical noise of the entire electronics (front-end module, amplifier, ADC) is very low and does even not play a significant role at very small signals around 1 mV.
- The instrument noise is thus entirely determined by the detector unit, i.e. by the sum of photon noise or quantum noise, respectively, and CPM noise. In Sect. 5, the quantum noise is calculated and compared with the observed noise.

4.2 Precision as function of measurement frequency

Next, the instrument noise is analyzed in dependence on the measurement frequency (which can be set using the three push buttons at the front plate, see Sect. 3).

The measurement speed of the sonde depends (i) on the integration time of the electronics and the data transfer rate of the serial interface which is both 50 Hz at maximum and (ii) on the exchange rate of sample air present between sensor disc and channel photomultiplier (CPM). This air exchange rate t is the ratio of the volume

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



V between sensor disc and CPM and the sample air flow f , i.e. $t = V/f$. V simply is the product of the sensor disc surface F ($=\pi \cdot d^2/4 = 0.8 \text{ cm}^2$) and the distance d between sensor disc and CPM ($d = 0.4 \text{ cm}$), i.e. $V = 0.32 \text{ cm}^3$. Already at very low flows of 1 l min^{-1} ($= 17 \text{ cm}^3 \text{ s}^{-1}$) the measurement volume is exchanged in 19 ms or with $\sim 50 \text{ Hz}$, respectively. Therefore, at usual measurement conditions the measurement frequency of the instrument itself is limited by the data transfer rate that is 50 Hz at most. However, for field use one has also to consider mixing that occurs in the inlet line which can make the effective response time much longer, see Sect. 4.5.

For quantifying the instrument precision in dependence on measurement frequency, data are analyzed that was recorded over a period of 100 min at a measurement frequency of 50 Hz and a constant O_3 mixing ratio of 45 ppbv, see Fig. 4.

The signal decreases exponentially, with an instantaneous sensitivity decrease of -30% per hour at measurement start and -1.5% per hour after 100 min (dashed line, right y-axis). This relaxation to (almost) constant sensitivity constitutes a typical feature of the sensor discs and is due to two equilibration processes:

1. the sensitivity dependence on the humidity of the sample air. Schurath et al. (1991) report a sensitivity increase of $\sim 11 \%$ per percent relative humidity for similar types of sensor discs. As the sample air delivered by the ozone generator is dry (relative humidity $< 1 \%$) and the water-repellence of the sensor disc high (see sensor disc preparation, Sect. 4), the fresh (pre-ozonized) sensor disc needs time to equilibrate with the humidity of the sample air.
2. the sensitivity dependence on the O_3 mixing ratio of the sample air. Fresh sensor discs are quite insensitive to O_3 and have to pre-activated with an O_3 dose of $100\text{--}500 \text{ ppbv} \times \text{h}$ (Speusser 1989, Muller et al., 2010, and references therein). Thereafter, still a weak and slowly varying dependence of the sensor sensitivity on the ozone mixing remains. That is, low ozone levels in sample air does result in weak re-activation and the consumption of active sites may outweigh until equilibrium is reached again at slightly lower sensitivity levels.

As is obvious from Fig. 4, both equilibration processes are slow and much slower than the atmospheric variations one is usually interested in, see Sect. 6.

The noise or precision, respectively, in the (50-Hz) dataset shown in Fig. 4 is evaluated by applying the Allan-variance method, see Fig. 5.

As described by Werle et al. (2004), the Allan-variance describes the precision ΔO_3 of a device as function of the time τ over which the data is integrated. The Allan-variance method is applied to five sub-datasets covering a measurement time of 20 min each. As shown in Fig. 5, all five distributions start at a precision $\sigma f \sim 0.4$ ppbv at a measurement frequency of 50 Hz. Towards integration times of up to ~ 10 s, the precision ΔO_3 roughly decreases with the square root of the number of considered data points and thus with the square root of the integration time τ his linear decrease in a log-log-plot with a slope of -0.5 describes the behavior for statistically independent measurements (white noise) described by a normal and Poisson distribution (dashed line).

At certain integration times, the so-called stability time, highest precision $O_{3,\min}$ is achieved. Then drifts, in the present case the sensitivity change of the sensor disc (see Fig. 4), starts to play a role and ΔO_3 increases again with longer integration times. As listed in the box in Fig. 5, highest precisions of 0.016–0.035 ppbv are achieved at stability times of 10–40 s, in which the latest sub-dataset (with the weakest drift, see Fig. 4) expectedly shows the longest stability time and thus highest precision. Note, that this precision is more than a factor of 10 better than the one of UV photometers which typically amounts to 0.5 ppbv at a measurement time of 10 s.

We conducted further laboratory tests, for instance recorded data with different measurement frequencies of 50, 25, 10, and 1 Hz over typically 1 h, calculated the noise over subsets of 100 measurement points each, and finally plotted the emerging noise distribution as probability density functions (PDFs, not shown). These PDFs illustrate how often a certain noise/precision occurred in these 1-h datasets. Indeed, we inferred exactly the same result, as already derived from Fig. 5.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Namely, at measurement frequencies of up to 0.1 Hz the data are Poisson distributed, i.e. the instrument indicates shot noise. In this case, the signal-to-noise ratio SNR increases with the square root of the number of considered data points N , i.e. $\text{SNR} = \sqrt{N}$, and the relative instrument noise decreases inverse proportionally with \sqrt{N} (Kingston, 1978). As further explained in Sect. 5, the observed shot noise is caused by the fluctuation of photons detected by the photomultiplier, i.e. by quantum noise or photon noise, respectively. In this case, the relative instrument noise $\Delta O_3/O_3$ simply decreases with the number of photons n falling onto the detector as follows:

$$\frac{\Delta O_3}{O_3} \propto \frac{1}{\sqrt{n}} \quad (1)$$

and as function of the number of considered data points N and the integration time τ :

$$\frac{\Delta O_3}{O_3} = \left(\frac{\Delta O_3}{O_3} \right)_{N=1} \cdot \frac{1}{\sqrt{N}} = \left(\frac{\Delta O_3}{O_3} \right)_{\tau=1\text{ s}} \cdot \frac{1}{\sqrt{\tau}} \quad (2)$$

being $(\Delta O_3/O_3)_{N=1}$ the relative noise at 50 Hz, $(\Delta O_3/O_3)_{\tau=1\text{ s}}$ the relative noise at $\tau = 1\text{ s}$ and τ in seconds. In Fig. 5 or at 45 ppbv absolute, $(\Delta O_3/O_3)_{N=1}$ amounts to $\sim 1.22\%$ ($=0.55\text{ ppbv}/45\text{ ppbv} \cdot 100\%$) and $(\Delta O_3/O_3)_{\tau=1\text{ s}}$ to $\sim 0.20\%$ ($=0.09\text{ ppbv}/45\text{ ppbv} \cdot 100\%$).

As elucidated in Sect. 5, the measured signal output can be converted in a number of photons that reached the detector during the measurement time. At the constant conditions during this laboratory test and the equivalence of integration time (lower x-axis) and number of analyzed photons (Eqs. 1 and 2), in Fig. 5 also the number of detected photons used to calculate the displayed standard deviation is indicated (upper x-axis). For instance, over an integration time of 1 s, altogether $3.9 \cdot 10^5$ photons were detected which result in a relative quantum noise of $1/\sqrt{390\,000} = 0.0016$ or 0.072 ppbv absolute ($= 0.0016 \cdot 45\text{ ppbv}$) and thus just the value indicated in Fig. 5.

4.3 Precision as function of ozone mixing ratio

Next the instrument precision is measured at O_3 mixing ratios between 3 and 300 ppbv, see Fig. 6. These measurements were done at ambient pressure (1000 hPa) at a flow of 2 l min^{-1} provided by the ozone generator. The measurement frequency was set to 10 Hz. Again a log-log diagram was chosen and again an extremely compact linear decrease of the 1-sigma noise with the ozone mixing ratio was observed, with a slope of -0.505 and a correlation coefficient $R^2 = 0.998$:

$$\frac{\Delta O_3}{O_3} = \left(\frac{\Delta O_3}{O_3} \right)_{O_3=1\text{ppbv}} \cdot \frac{1}{\sqrt{O_3}} \approx \frac{3.9\%}{\sqrt{O_3}} \quad (3)$$

with $(\Delta O_3/O_3)_{O_3=1\text{ppbv}}$ the relative noise at $O_3 = 1$ ppbv and O_3 the ozone mixing ratio in ppbv. Thus, $\Delta O_3/O_3$ is 1.22 % at $O_3 = 10$ ppbv and 0.39 % at $O_3 = 100$ ppbv (at a measurement frequency of 10 Hz).

The explanation for this decreasing noise with increasing O_3 mixing ratio (Eq. 3) is identical with the one brought forward for explaining Fig. 5 and the Eqs. (1) and (2). The instrument shows a linear response, i.e. the O_3 mixing ratio scales with the number of photons emitted by the sensor disc. Again (as in Fig. 5) the x-axis of Fig. 6 can thus be interpreted as number of detected/considered photons used to calculate the displayed relative instrument noise (upper x-axis). For example, at $O_3 = 100$ ppbv altogether $8.57 \cdot 10^4$ photons were detected which result in a relative quantum noise of $1/\sqrt{85\,700} = 0.34\%$ and thus just the value indicated in Fig. 6.

4.4 Signal as function of air sample flow

The sample air flow constitutes the last parameter that still influences the measurement signal and thus the measurement noise. Namely, the sample air flow determines the kinetics around the sensor disc and thus the afflux of O_3 molecules reacting on the sensor disc surface. Small sample air flows can e.g. result in an increasing depletion of O_3 molecules along the sensor disc as the exchange with undisturbed sample air is

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



so slow that the air layer just above the sensor disc is progressively diluted towards the end of the sensor disc. This kinetic limitation was found in earlier studies and results in the typical signal increase with increasing flow until a plateau is reached along which only a very weak dependence exists (Hodgeson et al., 1970; Speuser et al., 1989).

Hilsenrath and Kirschner (1980) described the signal S as function of the sample air flow F as follows:

$$S = A \cdot O_3 \cdot F \cdot \left(1 - e^{-B/F}\right) \quad (4)$$

with A and B empirical constants depending on sample disc sensitivity and sensor geometry. At low flows, S scales with the flow F ($S = A \cdot O_3 \cdot F$), and at high flows, S is independent on F ($S = A \cdot O_3 \cdot B$).

In Fig. 7, the signal intensity is displaced as function of the sample air flow.

The signal intensity strongly increases with the sample flow and reaches a plateau at flows of $>4.5 \text{ l min}^{-1}$. A fit using Eq. (4) can qualitatively describe the signal intensity but suggests too low signals below $\sim 1.2 \text{ l min}^{-1}$ and too high signals between 1.2 and 3.5 l min^{-1} .

4.5 Instrument response time

Besides the exchange rate of the air volume above the detector disc (which is only some ms, see Sect. 4.2), the instrument response time is determined by mixing within the air sample inlet tube. This mixing causes progressive attenuation of tracer (here O_3) fluctuations along the inlet line and is determined by (i) the inlet tube length and diameter, i.e. the transport time from the inlet tube tip to the sensor disc, and (ii) the flow regime in the inlet tube (laminar or turbulent).

For measuring the instrument response time, we conducted the following measurement sequence. Sample air having an O_3 mixing ratio of ~ 200 ppbv was guided via a solenoid PTFE valve and a (one connection to the laboratory open) T-fitting to the O_3 sonde. After a constant O_3 signal was reached, the solenoid valve was actuated after which the O_3 instrument immediately started to suck O_3 -poor laboratory air via

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

the T-fitting. We varied sampling flow and the length and diameter of the inlet tube, and analyzed the signal from the O₃ sonde as function of time. The measured O₃ signal for two different inlet tube lengths and sampling flows is depicted in Fig. 8. The air transport time to the sensor τ_{inlet} , defined here as the time between valve switch and initial O₃ signal change, amounted 0.45 s and 12.44 s for the two configurations. The τ_{90} time, defined as the time during which 90 % of the signal change occurred and which can be interpreted as the instrument response time, was 0.37 s and 1.62 s, respectively.

The same measurement sequence was repeated for sampling flows between 0.8 l min⁻¹ and 7.0 l min⁻¹ and inlet tubes having diameters of 1/4" and 3/8" and lengths between 0.11 m and 10 m, see results in Fig. 9a. As expected, the longer the air transport time in the inlet tube, either by increasing the inlet tube length or by decreasing the sampling flow, the larger is the instrument response time τ_{90} .

As documented by Taylor (1953, 1954) and Lenschow and Raupach (1991), this mixing along the axis *z* of a round inlet line can be described by the one-dimensional diffusion equation

$$\frac{\delta O_3}{\delta t} = K_z \cdot \frac{\delta^2 O_3}{\delta z^2} \quad (5)$$

with K_z the virtual diffusion coefficient being

$$K_z = \frac{r^2 \cdot u^2}{48 \cdot \nu_{O_3}} = \frac{r \cdot u \cdot Re}{96} \cdot Sc \quad \text{laminar flow Taylor et al. (1953)} \quad (6)$$

$$K_z = 1.01 \cdot \frac{r \cdot u}{Re^{0.125}} \quad \text{turbulent flow Lenschow and Raupach (1991)} \quad (7)$$

with r the tube inner radius, u the mean velocity of flow, ν_{O_3} the molecular diffusion coefficient of O₃, Re the Reynolds number being $2 \cdot r \cdot u / \nu_{\text{air}}$, ν_{air} the molecular diffusion coefficient of air and $Sc = \nu_{\text{air}} / \nu_{O_3} = 0.85$ the Schmidt number (Taylor et al., 2007).

As described in Taylor (1953, chapter 5, case B2) and Taylor (1954, chapter 10), for our case, namely the sudden change of concentration at time $t = 0$, the solution of the differential Eq. (5) describes the mean O_3 mixing ratio along the tube at a certain location z :

$$O_3(z) = A + B \cdot \operatorname{erf} \left(\frac{z - z_0}{\sqrt{k_z \cdot t}} \right) \quad (8)$$

with A the mean between the initial value $O_3^1 = O_3(t = 0)$ and end value $O_3^2 = O_3(t \gg 0)$, i.e. $A = (O_3^1 + O_3^2)/2$, B the half of the difference of the two O_3 values, i.e. $B = (O_3^1 - O_3^2)/2$, erf the well-known error function that goes from -1 to $+1$, and z_0 the location just at the middle of the transition, i.e. where $\operatorname{erf}(\dots) = 0$ and $O_3(z_0) = A = (O_3^1 + O_3^2)/2$. As we measure the O_3 mixing ratio as function of time (and not as function of z), in Eq. (8) we have to replace z by $t = z/u$, i.e.

$$O_3(t) = \frac{O_3^1 + O_3^2}{2} + \frac{O_3^1 - O_3^2}{2} \cdot \operatorname{erf} \left(\frac{t - t_0}{\sqrt{k_z \cdot t}} \cdot u \right) \quad (9)$$

All required numbers are known or measured so that Eq. (9) should describe the course from the initially high value O_3^1 to the low end value O_3^2 as function of time. In Fig. 8 the O_3 signal change suggested by Eq. (9) is indicated for both configurations as dashed line. In both cases the agreement with the measurement is quite good, although the theory suggests a slightly smoother transition especially for the short inlet tube and the higher flow.

Likewise in Fig. 9 solutions of the mixing model (Eq. 9) are displayed (grey and red areas) for inlet tube lengths of 0.11 m and 10 m and sample flows between 1.0 and 7.0 l min⁻¹, in Fig. 9a for 3/8" sampling tubes and in Fig. 9b for 1/4" sampling tubes. The upper border of the two areas represents the solution for laminar flow (K_z in Eq. 6) where a parabolic air speed distributions (with $u_{\max} = 2u$ in the tube center) leads to strong tracer attenuation along the tube. The lower border of the two areas represents

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Discussion Paper | Discussion Paper | Discussion Paper | Discussion Paper | Discussion Paper

the solution of Eq. (9) for turbulent flow (K_z in Eq. (7)) where $u_{\max} \sim 1.1 u$ (Lenschow and Raupach, 1991).

At first glance, one would expect that the solution for laminar flow should well approximate the observations, as the Reynolds number was ever well below the critical value of ~ 2300 , (Fig. 9a and b, upper x-axes) above which turbulence in straight tubes usually sets in. However, as demonstrated by Lenschow and Raupach (1991) each elbow and even soft bending in the tubing cause enhanced radial mixing in the tube and pushes the effective diffusion coefficient K_z towards the one for turbulent flow, even at Reynolds numbers below 1000.

Indeed our observations agree reasonably well with the solution for laminar flow at flows below $\sim 2.5 \text{ l min}^{-1}$ (see also the agreement in Fig. 8) and progressively deviate towards the solution for turbulent flow for higher sample flows. At low flows of below $1\text{--}3 \text{ l min}^{-1}$ (depending on tube diameter and length), the measured τ_{90} time exceeds the τ_{90} time calculated even for laminar flow. We believe that these longer τ_{90} times are due to some dead volumes, first between the valve and the T-fitting (outside the O_3 -sonde) and secondly within the PTFE detector block. Figure 9b indicates that with a short $1/4''$ sampling tube and the internal van, τ_{90} times of below 100 ms ($>10 \text{ Hz}$) can be reached, see further discussion in Sect. 6.2.

5 Instrument performance

In Sect. 4, we demonstrated that the instrument noise is (a) dominated by noise generated at the detector, i.e. the photomultiplier, and (b) decreases with the square root of the integration time τ and O_3 mixing ratio. In the following we will calculate the number of photons (photon flux) reaching the channel photomultiplier (CPM) and the relevant photon noise or quantum noise, respectively, and will compare this quantum noise with the noise actually measured.

The photon current I_{photon} can be derived from the current delivered by the channel photomultiplier I_{CPM} , the current amplification Z of the CPM (up to 10^8 , see CPM

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



datasheet), and the quantum efficiency Q_E of the CPM ($\sim 15\%$, see CPM datasheet):

$$I_{\text{photon}} = \frac{I_{\text{CPM}}}{Z \cdot Q_E} \quad (10)$$

The CPM current I_{CPM} is measured by the instrument as voltage U falling across the amplification resistance R ($2 \text{ M}\Omega$):

$$I_{\text{CPM}} = \frac{U}{R} = \frac{\frac{\text{counts}}{8388608} \cdot \frac{5 \text{ V}}{\text{PGA}}}{2 \text{ M}\Omega} \quad (11)$$

with counts voltage given by the 24-bit analog-digital-converter (ADC)

PGA internal digital amplification (set at the display between 1 and 64)

The photon current I_{photon} is generated by n photons each producing (with the quantum efficiency Q_E , see Eq. 10) one electron on the CPM detector surface:

$$I_{\text{photon}} = \frac{d(n_{\text{photon}} \cdot e^-)}{dt} = e^- \cdot \frac{dn_{\text{photon}}}{dt} \quad (12)$$

with e^- the elementary charge of $1.6 \cdot 10^{-19} \text{ C}$.

Thus, at usually once set CPM voltage U_{CPM} and digital amplification PGA, the number of photons falling onto the channel photomultiplier per second is:

$$\frac{dn_{\text{photon}}}{dt} = \frac{I_{\text{photon}}}{1.6 \cdot 10^{-19} \text{ C}} = \frac{\frac{\text{counts}}{8388608} \cdot \frac{5 \text{ V}}{\text{PGA}}}{2 \text{ M}\Omega \cdot Z \cdot Q_E \cdot 1.6 \cdot 10^{-19} \text{ C}} \quad (13)$$

At atmospheric pressure, an ozone mixing ratio of 45 ppbv, and typical values of $\text{PGA} = 4$ and $U_{\text{PM}} = 1700 \text{ V}$ ($Z = 441\,000$, see CPM datasheet), the instrument actually measured a signal of $\sim 75\,625$ counts and (according to Eqs. 10 and 11) a photo current I_{photon} of $\sim 85 \text{ fA}$, respectively. Equation (13) gives a photon flux of dn_{photon}/dt of $\sim 5.3 \cdot 10^5$ photons falling per seconds (cps) onto the detector (although only $Q_E = 15\%$ are detected).

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



The respective photon noise or quantum noise Δn_{photon} follows Poisson statistics (which approaches a normal distribution for large photon numbers) and thus is the square root of the number of detected photons:

$$\Delta n_{\text{photon}} = \sqrt{n_{\text{photon}}} \quad (14)$$

5 Finally, the relative quantum noise is

$$\frac{\Delta n_{\text{photon}}}{n_{\text{photon}}} = \frac{\sqrt{\text{cps}}}{\text{cps}} = \frac{1}{\sqrt{\text{cps}}} \quad (15)$$

However, the noise detectable by the CPM is higher, as its quantum efficiency Q_E is only $\sim 15\%$. This constitutes that only $\sim 15\%$ of the photons reaching the CPM are detected and the detectable quantum noise is:

$$\frac{\Delta n_{\text{photon}}}{n_{\text{photon}}} \Big|_{\text{CPM}} = \frac{\sqrt{\text{cps} \cdot Q_E}}{\text{cps} \cdot Q_E} = \frac{1}{\sqrt{\text{cps} \cdot Q_E}} \quad (16)$$

Inserting the number of photons reaching the detector in 1 s and 20 ms lead to a detectable quantum noise of:

$$\begin{aligned} \frac{\Delta n_{\text{photon}}}{n_{\text{photon}}} \Big|_{\text{CPM}} &= \frac{1}{\sqrt{2.6 \cdot 10^6 \cdot 0.15}} \approx 0.16\% \text{ at } 1 \text{ Hz} \\ &= \frac{1}{\sqrt{5.2 \cdot 10^4 \cdot 0.15}} \approx 1.13\% \text{ at } 50 \text{ Hz} \end{aligned} \quad (17)$$

15 As shown in Fig. 5, the noise actually measured is $\sim 0.18\%$ ($= 0.08 \text{ ppbv}/45 \text{ ppbv} \cdot 100\%$) at 1 Hz and $\sim 1.11\%$ ($= 0.5 \text{ ppbv}/45 \text{ ppbv} \cdot 100\%$) at 50 Hz, which exactly equals with the quantum noise.

As just shown, we are able to convert the measured signal into a number of detected photons and thus into a quantum noise which agrees with the instrument noise. This finding allows us to parameterize the absolute and the relative precision of the instrument as function of the O_3 mixing ratio and the set measurement frequency f , in the following with a typical sensitivity of our Coumarin-47 sensor discs:

$$\Delta \text{O}_3 \approx 0.01 \text{ ppbv} \cdot \sqrt{\text{O}_3} \cdot \sqrt{f} \quad (18)$$

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



$$\frac{\Delta O_3}{O_3} \approx 1\% \cdot \frac{\sqrt{f}}{\sqrt{O_3}} \quad (19)$$

This precision is plotted in Fig. 10.

Figure 10 nicely illustrates the power of the dry CI technique and specifically of the new instrument, namely its high measurement speed of up to 50 Hz at still very high precision. At a measurement frequency of 10 Hz and typical O₃ mixing ratios in the boundary layer of 10–100 ppbv, the relative precision amounts to 0.4–1.0 %. At an integration time of 10 s, the precision is another factor of 10 higher, i.e. 0.04–0.1 %, which is roughly a factor of 10 better than state-of-the-art UV photometers.

6 Field use, calibration, and recommendations

The strength of this new instrument is its outstanding precision at high measurement frequency of up to 50 Hz. It will play out its strength in particular for eddy-covariance flux measurements, usually at ground (Rummel et al., 2007) or ship (Bartbeau et al., 2010) and aircraft observations (Ulanovsky et al. 2001; Bradshaw et al., 2002; Zahn et al., 2002).

6.1 Recommendations concerning calibration

Essential for both applications is the frequent calibration with an absolute measuring device, usually an UV-photometer. Indeed, reliable and accurate calibration is an eminently important issue, which in our opinion requires a more stringent approach than we could often find in the literature. We raise three arguments that substantiate our concern:

1. Muller et al. (2010) addressed the calibration of dry CI instruments in detail. They suggest inferring the instrument sensitivity based on 15 or 30 min averages, in which three approaches differing in the way of data treatment are analyzed (e.g. if

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



instrument offsets are considered or not). These three approaches led to considerably different sensitivities and thus inferred ozone fluxes.

2. Guesten et al. (1992) observed a considerable enhancement of the sensor sensitivity by $\sim 40\%$ after abrupt changes of the absolute humidity from 0 to 1%. The τ -90 response time was estimated to ~ 22 s. Although such sudden humidity changes do not occur in the atmosphere (at least not on stationary measurement platforms), this strong H_2O cross-sensitivity might introduce a significant source of uncertainty at too rare calibration sequences.
3. We collected considerable information on the sensitivity of our dry CI instrument deployed during ~ 300 long-distance flights onboard the CARIBIC passenger aircraft to various destinations worldwide, during all seasons, and at a wide range of humidities. The data processing after flight runs as follows: (1) Accurate O_3 mixing ratios are inferred from a home-made UV-photometer at a time resolution of 4 sec and an accuracy of ~ 0.5 ppbv. (2) From this O_3 dataset averaged data ($\text{O}_{3,\text{mean}}$) are generated with a set integration time of usually 15 s. (3) Over the same time period pressure-corrected averaged signals of the O_3 -sonde are retrieved (V_{mean}/p), the (extremely small) background signal ($V_{\text{mean},0}/p$) subtracted, and the instantaneous O_3 -sonde sensitivity calculated: $(V_{\text{mean}} - V_{\text{mean},0})/p/\text{O}_{3,\text{mean}}$. (4) Using this calibration file, 10 Hz - ozone data are inferred from the (10 Hz) data recorded from the O_3 -sonde.

One typical example is shown in Fig. 6. Here, the (4 s) UV photometer derived O_3 data are compared with the O_3 -sonde data likewise averaged over 4s (or 40 data points). Plotted is the frequency a certain difference of the two instruments occurred during the flight (in 1 % difference steps).

It is obvious that the high-frequency calibration works much better. The mean difference is $-0.1 \pm 2.6\%$ compared to $-4.9 \pm 12.4\%$ for the 15min calibration. Moreover, large deviations of -25 to -40% occurred during ascent and descent where fast

changes of the humidity altered the sensitivity of the sensor disc, so that a 15min calibration lags behind this fast sensitivity change.

Based on these three arguments we strongly recommend to apply ever a continuous calibration based on an UV-photometer with a time constant of <1 min (for aircraft application) and ~ 5 min (for eddy co-variance observations). We moreover recommend to measure the background signal (zero-point) of the O_3 -sonde (which is usually very small and constant, as determined by internal signal amplifiers) about once a week (simply by closing air inlet and outlet) and to subtract this zero signal from the measured signal before a calibration procedure is applied.

6.2 Recommendations concerning sample air flow

As described in Sect. 4.5, the effective response time of the device is solely influenced by mixing occurring in the sampling tube between tube tip and sensor disc. The instrument is equipped with a high-efficient van (Sect. 2) which drives sample air through the device of up to 5 l min^{-1} depending on the length of the sampling tube. As indicated in Fig. 9b and if considering that some unavoidable dead volumes in the test set-up have slightly increased the measured τ_{90} times (see Sect. 4.5), short $1/4''$ sampling tubes of ~ 10 cm (just to avoid contamination from the instrument itself) allow τ_{90} response times of below 100 ms (>10 Hz). When longer sampling tubes have to be used and/or shorter response times are required, we recommend connecting an additional external pump that increases the sample flow and the degree of turbulence within the sampling tube. By applying Eq. (9), the user can easily estimate the response time for arbitrary sample tube geometries (inner diameter, length) and sampling flows. An upcoming redesign will further optimize the detector block concerning short response times by avoiding dead volumes and sudden changes of the cross-section of the sampling channel.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

6.3 Recommendations concerning electric connection and data transfer

Its outstanding precision of usually $<1\%$ at 10 Hz (see Fig. 6) makes the O_3 -sonde potentially susceptible to any improper electrical connection such as ground loops or data cables being too long or having too high capacities. We thus recommend (i) powering the instrument via a passably clean power supply, (ii) grounding the chassis only if the RS-232 interface is not connected, and (iii) most importantly, collecting data – if possible – solely via the (digital) RS-232 interface which is much more robust against electro-magnetic interferences than the analog output.

7 Conclusions

The O_3 -sonde characterized here shows an outstanding performance and makes it to one of the fastest device for measuring atmospheric trace gases. It is furthermore lightweight, robust, and does not contain hazardous materials, which allows facile deployment even on high towers, ships, and aircraft.

The instrument precision is quantum noise limited. Significant further improvement of the instrument can thus only be achieved by increasing the number of photons emitted by the sensor discs. This in turn may be achieved by two improvements: 1. the production of more sensitive sensor discs and 2. higher turbulence around the sensor disc so that more O_3 molecules reach the detector and trigger more photons. Both “improvements” however lead to higher consumption of activatable sites on the sensor disc that shortens its lifetime by the same factor the sensitivity was enlarged.

One current problem actually constitutes the supply of sensor discs. To our knowledge the sole supplier to date is Bagus Consulting (Speyer, Germany) whose sensor discs currently show high sensitivity but fast aging that limits the use to 1–2 days. This characteristic can slightly be adopted to longer durability but with lower sensitivity (H. Güsten, personal communication, 2011). Our ~ 10 -yr old GFAS sensor discs are fairly sensitive and very durable and are typically exchanged after 12 flights (altogether ~ 150 h) or an O_3 -dose of $\sim 15\,000$ ppbv·h. However, the company GFAS

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

crashed and the disposal of sensor discs with it. We could initiate the renewal of the research on sensor discs at the Max-Planck-Institute of Chemistry, Mainz (in the group of F. Meixner) in collaboration with the university Mainz (in the group of T. Hoffmann). Objective is the development of sensitive and durable sensor discs by testing different dyes, sensor disc materials, and hydrophobization sequences. It can currently not be foreseen when these new types of sensor discs will be commercially available.

Acknowledgements. We acknowledge the support of the Deutsche Forschungsgemeinschaft and of the European Commission through the GEOmon (Global Earth Observation and Monitoring) Integrated Project under the 6th Framework Program (contract FP6-2005-Global-4-036677).

References

- Bariteau, L., Helmig, D., Fairall, C. W., Hare, J. E., Hueber, J., and Lang, E. K.: Determination of oceanic ozone deposition by ship-borne eddy covariance flux measurements, *Atmos. Meas. Tech.*, 3, 441–455, doi:10.5194/amt-3-441-2010, 2010.
- Bradshaw, N. G., Vaughan, G., Busen, R., Garcelon, S., Jones, R., Gardiner, T. T., and Hacker, J.: Tracer filamentation generated by small-scale Rossby wave breaking in the lower stratosphere, *J. Geophys. Res.*, 107, 4689, doi:10.1029/2002JD002086, 2002.
- Coyle, M.: The gaseous exchange of ozone at terrestrial surfaces: non-stomatal deposition to grassland, PhD thesis, University of Edinburgh, UK, 2005.
- Guesten, H. and Heinrich, G., On-line measurements of ozone surface fluxes: Part I. Methodology and instrumentation, *Atmos. Environ.*, 30, 897–909, 1996.
- Guesten, H., Heinrich, G., Schmidt, R. W. H., and Schurath, U.: A novel ozone sensor for direct flux measurements, *J. Atmos. Chem.*, 14, 73–84, 1992.
- Hegglin, M. I., Brunner, D., Peter, T., Hoer, P., Fischer, H., Staehelin, J., Krebsbach, M., Schiller, C., Parchatka, U., and Weers, U.: Measurements of NO, NO_y, N₂O, and O₃ during SPURT: implications for transport and chemistry in the lowermost stratosphere, *Atmos. Chem. Phys.*, 6, 1331–1350, doi:10.5194/acp-6-1331-2006, 2006.
- Hilsenrath, E. and Kirschner, P. T.: Recent assessment of the performance and accuracy of

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

a chemiluminescent rocket sonde for upper atmospheric ozone measurements, *Rev. Sci. Instrum.*, 51, 1381–1389, 1980.

Hodgeson, J. A., Krost, K. J., O’Keeffe, A. E., and Stevens, R. K.: Chemiluminescent measurement of atmospheric ozone: response characteristics and operating variables, *Anal. Chem.*, 42, 1795–1802, 1970.

IPCC: Fourth Assessment Report (AR4), *Climate Change*, 2007.

Kingston, R. H.: Detection of optical and infrared radiation, *Springer Series in Optical Sciences*, 10, 1978.

Kurpius, M. R., McKay, M., and Goldstein, A. H.: Annual ozone deposition to a Sierra Nevada ponderosa pine plantation, *Atmos. Environ.* 36, 4503–4515, 2002.

Levy, H.: Normal atmosphere: Large radical and formaldehyde concentrations predicted, *Science*, 173, 141–143, 1971.

Lenschow, D. H. and Raupach, M. R.: The attenuation of fluctuations in scalar concentrations through sampling tubes, *J. Geophys. Res.*, 96, 15259–15268, 1991.

Li, J., Li, Q., Dyke, J. V., and Dasgupta, P. K.: Atmospheric ozone measurement with an inexpensive and fully automated porous tube collector-colorimeter, *Talanta*, 74, 958–964, 2008.

McKendry, I. G., Steyn, D. G., O’Kane, S., Zawar-Reza, P., and Heuff, D.: Lower tropospheric ozone measurements by light aircraft equipped with chemiluminescent sonde, *J. Atmos. Oceanic Technol.*, 15, 136–143, 1998.

Muller, J. B. A., Percival, C. J., Gallagher, M. W., Fowler, D., Coyle, M., and Nemitz, E.: Sources of uncertainty in eddy covariance ozone flux measurements made by dry chemiluminescence fast response analysers, *Atmos. Meas. Tech.*, 3, 163–176, doi:10.5194/amt-3-163-2010, 2010.

Ray, J. D., Stedman, D. H., and Wendell, G. J.: Fast chemiluminescent method for measurement of ambient ozone, *Anal. Chem.*, 58, 598–600, 1986.

Regener, V.: On a sensitive method for the recording of atmospheric ozone, *J. Geophys. Res.*, 65, 3975–3977, 1960.

Ridley, B. A., Grahek, F. E., and Walega, J. G.: A small, high-sensitivity, medium-response ozone, detector suitable for measurements from light aircraft, *J. Atmos. Oceanic Technol.*, 9, 142–148, 1992.

Rummel, U., Ammann, C., Kirkman, G. A., Moura, M. A. L., Foken, T., Andreae, M. O., and Meixner, F. X.: Seasonal variation of ozone deposition to a tropical rain forest in southwest

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

Amazonia, *Atmos. Chem. Phys.*, 7, 5415–5435, doi:10.5194/acp-7-5415-2007, 2007.

Sahand, S.: Optimierung der Oberflächen-Chemilumineszenz für Ozonmessungen in der Atmosphäre, Ph.D. work, University Bonn, Germany, 1989 (in German).

Schurath, U., Speuser, W., and Schmidt, R.: Principle and application of a fast sensor for atmospheric ozone, *Fresenius, J. Anal. Chem.*, 340, 544–547, 1991.

Speuser, W.: Eine Chemilumineszenzsonde zur Messung von Ozonprofilen in der Atmosphäre, Ph.D. work, University Bonn, Germany, 1989 (in German).

Takayanagi, T., Su, X.-L., Dasgupta, P. K., Martinelango, K., Li, G., Al-Horr, R. S., and Shaw, R. W.: Chemiluminometric measurement of atmospheric ozone with photoactivated chromotropic acid, *Anal. Chem.*, 75, 5916–5925, 2003.

Taylor, G. I.: Dispersion of soluble matter in solvent flowing slowly through a tube, *Proc. R. Soc. London, Ser. A*, 219, 186–203, 1953.

Taylor, G. I.: The dispersion of matter in turbulent flow through a pipe, *Proc. R. Soc. London, Ser. A*, 223, 446–468, 1954.

Vömel, H. and Diaz, K.: Ozone sonde cell current measurements and implications for observations of near-zero ozone concentrations in the tropical upper troposphere, *Atmos. Meas. Tech.*, 3, 495–505, doi:10.5194/amt-3-495-2010, 2010.

Weinheimer, A. J.: Chemical methods: chemiluminescence, chemical amplification, electrochemistry, and derivatization, in: “Analytical techniques for atmospheric measurement”, edited by: Heard, D. W., Blackwell Publishing, 2006.

Williams, E. J., Fehsenfeld, F. C., Jobson, B. T., Kuster, W. C., Goldan, P. D., Stutz, J., and McClenny, W. A.: Comparison of ultraviolet absorbance, chemiluminescence, and DOAS instruments for ambient ozone monitoring, *Environ. Sci. Technol.*, 40, 5755–5762, 2006.

World Meteorological Organisation. Scientific Assessment of Ozone Depletion: 2006, Rep. 50, Global Ozone Research and Monitoring Project, Geneva, 2007.

Ulanovsky, A. E., Yushkov, V. A., Sitnikov, N. M., and Ravengnani, F.: The FOZAN-II fast-response chemiluminescent airborne ozone analyzer, *Instr. Experim. Techn.*, 44, 249–256, 2001.

Werle, P., Mazzinghi, P., D’Amato, F., Rosa, M. D., Maurer, K., and Slemr, F.: *Spectrochim. Acta, Review: Signal processing and calibration procedures for in-situ diode-laser absorption spectroscopy, Part A*, 60, 1685–1705, 2004.

Yushkov, V., Oulanovsky, A., Lechenuk, N., Roudakov, I., Arshinov, K., Tikhonov, F., Stefanutti, L., Ravegnani, F., Bonafe, U., and Georgiadis, T.: A chemiluminescent analyzer for strato-

spheric measurements of the ozone concentration (FOZAN), J. Atmos. Ocean. Technol., 16, 1345–1350, 1999.

5 Zahn, A., Brenninkmeijer, C. A. M., Crutzen, P. J., Parrish, D. D., Sueper, D., Heinrich, G., Güsten, H., Fischer, H., Hermann, M., and Heintzenberg, J.: Electrical discharge source for tropospheric “ozone-rich transients”, J. Geophys. Res., 107, 4638, doi:10.1029/2002JD002345, 2002.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion






Fig. 1. Mechanical setup. The right sketch indicates the entire O_3 instrument integrated into a standard 3 HU, half 19" rack mount insert (top plate and back board removed). The front plate is equipped with inlet/outlet fittings, power and data interface socket (RS232, USB), fuse, display, three push buttons, and the flap behind which the quick-release knob (sensor disc holder) is located. The left exposition details the two compartments where the photomultiplier and the electronics are housed. The sample air channel and the arrangement around the sensor disc are displayed at the top. On request, the instrument is delivered within a splash guard cabinet.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page	
Abstract	Introduction
Conclusions	References
Tables	Figures
⏪	⏩
◀	▶
Back	Close
Full Screen / Esc	
Printer-friendly Version	
Interactive Discussion	

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.




Fig. 2. Function blocks of the electronics. Compartment a houses the photomultiplier and the high voltage supply, and compartment b the control unit and data acquisition system, see Fig. 1.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.




Fig. 3. Absolute signal (**graph a**) and relative noise (**graph b**) in dependence on the photomultiplier voltage at an O_3 mixing ratio of 45 ppbv and a measurement frequency of 10 Hz.

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.




Fig. 4. Normalized signal (left y-axis) along a period of 100 min. Light grey: normalized raw data recorded at 50 Hz. Dark grey: data averaged over 1 s. White straight line: exponential fit of the normalized data. Black dashed line: derivative of the exponential fit which gives the signal change in percent per hour (right y-axis).

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.




Fig. 5. Allan plot for five sub-datasets covering 20 min of measurement each. The upper x-axis indicates the number of detected photons considered to derive the plotted standard deviation. For a more detailed explanation, see text. The box lists the highest precision $O_{3,\min}$ (in ppbv) achieved at the stability time τ_{\min} (in seconds) of each sub-dataset.

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.




Fig. 6. Relative noise or measurement precision, respectively, as function of ozone mixing ratio between 3 and 300 ppbv at a measurement frequency of 10 Hz. The upper x-axis indicates the number of detected photons considered to derive the plotted relative noise.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.




Fig. 7. Signal intensity as function of the air sample flow (in normal liter per minute). Dots: measured data. Dashed line: fit using Eq. (4).

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)




Fig. 8. Instrument response in dependence on sampling flow and length of the 3/8'' inlet tube. Signal response for two configurations; red: inlet tube length of $\Delta = 0.11$ m and sample flow of 2.51 min^{-1} ; blue: inlet tube length of $\Delta = 10$ m and sample flow of 1.51 min^{-1} . τ_{inlet} is the total air transport time in the inlet tube having a total length of 0.30 m internal tube length within the instrument plus Δ additional (external) tube length (in total: 0.30 m + Δ). τ_{90} is the time span along which 90% of the total signal change occurred. Dashed black lines indicate solutions of the diffusion equation for laminar flow, see Eq. (9).

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures

◀ ▶

◀ ▶

Back Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion




Fig. 9. Measured τ_{90} times for inlet tube lengths Δ between 0.11 m and 10 m and air sample flows between 0.8 l min^{-1} and 7.0 l min^{-1} . Top: for $3/8''$ (ID = 6 mm) sampling tube. Bottom: for $1/4''$ (ID = 4 mm) sampling tube. The upper x-axes indicate the Reynolds number. The grey and red areas are solutions of the diffusion Eq. (8), the upper border for pure laminar flow, the lower border for turbulent flow, in grey for an inlet tube length of $\Delta = 10 \text{ m}$, and in red of $\Delta = 0.11 \text{ m}$.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures

◀ ▶

◀ ▶

Back Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

**Chemiluminescence
ozone detector for
eddy flux and
airborne application**

A. Zahn et al.




Fig. 10. Absolute precision (**graph a**) and relative precision (**graph b**) as function of ozone mixing ratio and measurement frequency.

Chemiluminescence ozone detector for eddy flux and airborne application

A. Zahn et al.




Fig. 11. Plot illustrating the quality of the calibration of our dry CI sensor with a UV photometer during a CARIBIC flight of ~ 10 h from Frankfurt (Germany) to Vancouver (Canada) on 24 September 2009. x-axis: difference (in %) of the two instruments. y-axis: frequency a difference occurred. The total number of 4s-data points is 8783. For the calibration two different integration times are applied: 15 s (full dots), 15 min (light dots). The mean difference and standard deviation of the two datasets are given in the boxes. The mean O_3 mixing ratio during flight was 204 ppbv, the minimum 29 ppbv, and the maximum 351 ppbv.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion